

To: Wall, Dan[wall.dan@epa.gov];
M.Blanchard@WestonSolutions.com[M.Blanchard@WestonSolutions.com]
From: Chapin, Thomas
Sent: Sun 8/9/2015 5:45:07 PM
Subject: Re: FW: GKM SAP-QAPP initial draft

Hi Mark and Dan,

Here are the corrections for the MiniSipper component to the SAP in Worksheet 22 Field Equipment Calibration, Maintenance, Testing, and Inspection. Corrections in to table on page 39

MiniSipper	NA	Replace batteries and filters and clean pump tubing and sample coil	Operational equipment check	Visually inspect for obvious defects, broken parts, or cleanliness	Prior to use	Equipment operational	Repair/replace as needed	USGS	WAM/COR	Instrumen Specific
------------	----	---	-----------------------------	--	--------------	-----------------------	--------------------------	------	---------	--------------------

On Sun, Aug 9, 2015 at 10:07 AM, Wall, Dan <wall.dan@epa.gov> wrote:

Please provide comments to mark blanchard ASAP. I recommend you focus on parts of your particular interest/activities. thanks

--

Thanks,

Thomas

Thomas Chapin, Ph.D.
Research Chemist
U.S. Geological Survey
Crustal Geophysics and Geochemistry Science Center
Box 25046, Mailstop 964D
Denver Federal Center
Denver, CO 80225

phone: 303-236-5795
fax: 303-236-3200
email: tchapin@usgs.gov
